

GOVERNMENT OF MAHARASHTRA
GOVERNMENT MEDICAL COLLEGE & HOSPITAL, MIRAJ
Pandharpur Road, Miraj Pin Code 416 410 Dist SANGLI

Phone No : STD Code No 0233- 2232100 to 2232109 (PBX)

Dean-(Chamber) 0233-231158

Fax No. 0233-233-2231959

GMC&HM/SS/2898/2018

Date- 5 / 12 / 2018

On line Quotation


T0,

You are requested to furnish item wise rates on or before 7 - 12 - 2018. For the material mentioned below. The general terms and conditions will be as follows.

1. Rates should be F.O.R. destination, inclusive of all taxes and duties and valid for at least 6/12 months.
2. Details, Specifications, Literature, Catalogues etc. should be attached along with the quotation.
3. Name of the manufacturer would be mentioned by Distributors, wholesalers without fail.
4. Exact delivery period would be mentioned without fail.
5. Firm should have valid licenses.
6. The sailed quotations should be addressed to the Dean, Govt. Medical College & Hospital, Miraj

7-Octri Exemption

Sr. No	Item	Specification	Rate
1	Micropipette 50-200 Adjustable Volume	attached	Each no
2	Micropipette 100 Fixed Volume	attached	Each no
3	Micropipette 200 Fixed Volume	Attached	Each no
4	Incubator	Attached	Each no
5	K2 EDTA Tube		Each no
6	Reticuleocyte stain Kit		Each no


DEAN,
Govt. Medical College & Hospital, Miraj

MICROPIPETTE—FIXED VOLUME - SPECIFICATIONS

2

3

MICROPIPETTE— FIXED VOLUME	VOLUME	INACCURACY+/-	IMPRECISION +/-
	100 <i>ul</i>	0.5	0.3
	200 <i>ul</i>	0.4	0.2

MICROPIPETTE ADJUSTABLE VOLUME - SPECIFICATIONS

1

MICROPIPETTE ADJUSTABLE VOLUME	VOLUME	INCREMENT	INACCURACY+/-	IMPRECISION +/-
	50-200 <i>ul</i>	1	1	0.15-0.4

5

INCUBATOR

TECHNICAL SPECIFICATION

- Dimensions : 58.0 cm x 54.0 cm x 58.0 cm in W x D x H cm
- Volume – 150 lt.
- Capacity- 9 lt
- Digital Display
- Temperature alarm
- Accuracy - $\pm 0.5^{\circ}\text{C}$
- Temperature Range 0.1°C up to 120°C Uniformity $\pm 0.25^{\circ}\text{C}$ @ 37°C
- Stability $\pm 0.2^{\circ}\text{C}$ from 37 to 100°C
- Electrical 115 VAC, 50/60 Hz, 380 W 230 VAC, 50/60 Hz, 380 W (CE)